

## Spectral Gamma-Ray Borehole Log Data Report

Page 1 of 2

Log Event A

# Borehole 30-03-05

# **Borehole Information**

**N-Coord**: 42,820 **W-Coord**: 48,165 **TOC** Elevation: 645.00

Water Level, ft : Date Drilled :  $\frac{7/31/1974}{}$ 

#### **Casing Record**

Type: Steel-welded Thickness: 0.280 ID, in.: 6

Top Depth, ft. :  $\underline{0}$  Bottom Depth, ft. :  $\underline{100}$ 

#### **Borehole Notes:**

This borehole was drilled in July 1974 to a depth of 100 ft with 6-in. casing. The casing thickness is presumed to be 0.280 in., on the basis of the published thickness for schedule-40, 6-in. steel tubing. No information concerning grouting or perforations was found; therefore, it is assumed that the borehole was not grouted or perforated. The top of the casing, which is the zero reference for the SGLS, is even with the ground surface.

## **Equipment Information**

 Logging System :
 1
 Detector Type :
 HPGe
 Detector Efficiency:
 35.0 %

 Calibration Date :
 10/1996
 Calibration Reference :
 GJO-HAN-13
 Logging Procedure : P-GJPO-1783

## Log Run Information

Log Run Number : 1 Log Run Date : 04/14/1997 Logging Engineer: Alan Pearson

Start Depth, ft.:  $\underline{0.0}$  Counting Time, sec.:  $\underline{100}$  L/R:  $\underline{L}$  Shield:  $\underline{N}$  Finish Depth, ft.:  $\underline{34.0}$  MSA Interval, ft.:  $\underline{0.5}$  Log Speed, ft/min.:  $\underline{n/a}$ 

Log Run Number: 2 Log Run Date: 04/15/1997 Logging Engineer: Alan Pearson

Start Depth, ft.:  $\underline{100.0}$  Counting Time, sec.:  $\underline{100}$  L/R:  $\underline{L}$  Shield:  $\underline{N}$  Finish Depth, ft.:  $\underline{33.0}$  MSA Interval, ft.:  $\underline{0.5}$  Log Speed, ft/min.:  $\underline{n/a}$ 



## Spectral Gamma-Ray Borehole Log Data Report

Page 2 of 2

Borehole 30-03-05

Log Event A

# **Analysis Information**

Analyst: D.L. Parker

Data Processing Reference : P-GJPO-1787 Analysis Date : 05/06/1997

#### **Analysis Notes:**

This borehole was logged by the SGLS in two log runs. The pre- and post-survey field verification spectra met the acceptance criteria established for the peak shape and detector efficiency, confirming that the SGLS was operating within specifications. The energy calibration and peak-shape calibration from these spectra were used to establish the channel-to-energy parameters used in processing the spectra acquired during the logging operation. There was some gain drift and it was necessary to adjust the established channel-to-energy parameters during processing of log data to maintain proper peak identification.

Casing correction factors for a 0.280-in.-thick steel casing were applied during analysis.

The only man-made radionuclide detected in this borehole was Cs-137. The presence of Cs-137 was measured almost continuously from the ground surface to 43 ft, continuously from 54.5 to 76 ft, and near the bottom of the borehole.

The K-40 concentrations increase gradually from 39 to 48 ft and remain relatively constant between 48 and 84 ft. The K-40 concentrations increase slightly at 84 ft.

Additional information and interpretations of log data are included in the main body of the Tank Summary Data Report for tank C-103.

### Log Plot Notes:

Separate log plots show the man-made and the naturally occurring radionuclides. The natural radionuclides can be used for lithology interpretations. The headings of the plots identify the specific gamma rays used to calculate the concentrations. Uncertainty bars on the plots show the statistical uncertainties for the measurements as 95-percent confidence intervals. Open circles on the plots give the MDL. The MDL of a radionuclide represents the lowest concentration at which positive identification of a gamma-ray peak is statistically defensible.

A combination plot includes the man-made and natural radionuclides, the total gamma derived from the spectral data, and the Tank Farms gross gamma log. The gross gamma plot displays the latest available digital data. No attempt has been made to adjust the depths of the gross gamma logs to coincide with the SGLS data.

A comparison plot is also provided showing the Cs-137 and Co-60 concentrations determined from the SGLS and those determined from the Radionuclide Logging System (RLS) in 1994.

A plot of representative historical gross gamma-ray logs from 1975 to 1982 is also included. The headings of the plots identify the date on which the data in the plots were gathered.